

DIGITAL AUTOMATIC LEAK TEST APPARATUS



VLT-A (300)

Advanced version of Model : LTA-300.

- Complete automatic system.
- Meets USP specifications.
- Display Indicates Set and Actual VACUUM LEVEL.
- VACUUM LEVEL can set as per requirement up-to-600 mm Hg.
- Settable Retention Time, Start delay Time and Vacuum hold Time.
- Automatic release of Vacuum pressure. Incorporated with inbuilt diaphragm (oil free) vacuum pump.
- Parallel port interface for dot

matrix printer.

- 20 x 4 alphanumeric informative LCD screen.
- 4 x 4 matrix self explanatory key board.

SPECIFICATIONS:

Desiccator : 300mm Polycarbonate vacuum desiccator

Instrument Body : SS 304

Instrument Dimensions : 290 mm x 470 mm
x 260 mm

Power Supply : 230 VAc, 50/60 Hz

The same instrument can be supplied with 150mm Or 250mm desiccator.